



BF-1TH190

Therapeutic bronchoscope with superb HDTV image quality and a 2.8 mm instrument channel.











HDTV


Superb HDTV images enable more detailed observation of bronchial surfaces.

NBI (Narrow Band Imaging)


NBI is an optical image enhancement technology that improves the visualisation of vessels on the mucosal surface. NBI is now 50% brighter compared to previous models.



Electronic magnification of $1.2\times$ and $1.5\times$ is now possible, allowing closer observation.



HDTV



NBI

Insertion tube rotation function

The insertion tube can be rotated left or right up to 120 degrees by simply turning a ring on the control section of the scope. This supports easier operation and smoother insertion with less operator fatigue.



Waterproof One-touch Connector

A new connector design minimises the effort required for set-up prior to and in between cases. In addition, it is fully submersible without a water-resistant cap and eliminates the associated risk of an expensive repair due to accidental immersion.



Ergonomic scope cable direction

The angle of direction of the universal cord has been improved so that the scope cable does not hang limp, even if the video system is located behind the bronchoscopist.

| C | :4 | | |
|-----|-------|-------|-----|
| -50 | ecifi | ICATI | ons |
| | | | |

| • | | | |
|-------------------------------------|--|----------------------------|--|
| Optical system | Field of view | 120° | |
| | Direction of view | Forward viewing | |
| | Depth of field | 3–100 mm | |
| | Image quality | HDTV | |
| Insertion section | Distal end outer diameter | 6.2 mm | |
| | Distal end enlarged | | |
| | Instrument channel outlet | Up Light-guide lens | |
| | Right | Left | |
| | | Objective lens | |
| | | Down | |
| | Insertion tube outer diameter | 6.0 mm | |
| | Working length | 600 mm | |
| | Insertion tube rotation function | n Yes | |
| Instrument channel | Channel inner diameter | 2.8 mm | |
| | Minimum visible distance | 3.0 mm from the distal end | |
| | Direction from which endotherapy accessories | | |
| enter and exit the endoscopic image | | | |



| Bending section | Angulation range | Up 180 ° — — — — — — — — — — — — — — — — — — |
|---|------------------------------------|--|
| Electrocautery instrument compatibility | Yes | |
| Laser compatibility | Nd: YAG, 810 nm diode | |
| Compatible | Video system center OLYMPUS CV-190 | |
| EVIS EXERA system | Xenon light source OLYMPUS CLV-190 | |

Specifications, design and accessories are subject to change without any notice or obligation on the part of the manufacturer.

